



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/672,137	09/26/2003	Robert A. Kruger	OPTO/11	6106
26875 7590 02/05/2008 WOOD, HERRON & EVANS, LLP 2700 CAREW TOWER 441 VINE STREET CINCINNATI, OH 45202			EXAMINER RAMIREZ, JOHN FERNANDO	
			ART UNIT 3737	PAPER NUMBER
			MAIL DATE 02/05/2008	DELIVERY MODE PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

## Office Action Summary

Application No.

10/672,137

Applicant(s)

KRUGER, ROBERT A.

Examiner

John F. Ramirez

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 13 November 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☐ Claim(s) \_\_\_\_\_ is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 10-22 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Claim Rejections - 35 USC § 112***

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 16-22 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. It is unclear as to how is the radiant energy delivered to the tissue and received from the tissue. It is also unclear as to how is the ultrasound beam delivered to the tissue and received. Therefore, it is unclear how is the image being generated from one or both signals.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-11, 14, 16, 18 and 21 are rejected under 35 U.S.C. 102(b) as being anticipated by Wang (US 6,567,688).

Wang discloses a thermoacoustic tomography (TACT) imaging system for imaging structures of tissue (abstract), comprising: a detector transducer array (see figure 2, element 26,); an acoustic receiver coupled to the detector array for receiving

acoustic signals generated in response to radiant energy delivered to the tissue (see abstract, col.6, lines 23-36) and detected by the detector array ; an ultrasound receiver coupled to the detector array for receiving echoes from an ultrasonic beam delivered into the tissue (col. 15, lines 7-21, fig. 9); and an ultrasound beam coupled to the detector array for causing the array to generate the ultrasonic beam (col. 15, lines 7-21, fig. 9).

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which the subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 12-13, 15, 17, 19-20 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Wang (US 6,567,688). Wang does not disclose specifically that the microwave-induced thermoacoustic system has a receiver and a beam steering circuit coupled to and controlled by an ultrasound imaging system and the images are presented on display overlaid in spacial registration or in comparison with each other. However, Wang discloses an alternative system that can be used to obtain ultrasonogram images and thermoacoustic images sequentially (see col. 14, lines 65-67, col. 15 lines 1-7). It would have been obvious to one of ordinary skill in the art at the time of the invention to include a receiver and a beam steering circuit coupled to and

controlled by an ultrasound system and to simultaneously displaying both images in order to enhance diagnostic information available to the physician.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John F. Ramirez whose telephone number is (571) 272-8685. The examiner can normally be reached on (Mon-Fri) 7:00 - 3:30 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian L. Casler can be reached on (571) 272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number:  
10/672,137  
Art Unit: 3737

Page 5

A handwritten signature in black ink, appearing to read "B. L. Casler", positioned above the printed name.

**BRIAN L. CASLER**  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 3700